

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020221\
 Data File : VW020237.D
 Acq On : 02 Feb 2021 14:57
 Operator : SY/MD
 Sample : M1258-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4Z4

Quant Time: Feb 03 02:29:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 03 02:27:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	759224	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.855	117	742904	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.254	152	377399	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	231526	43.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.82%
7) Chloroethane-d5	1.570	69	203768	50.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.46%
11) 1,1-Dichloroethene-d2	2.107	63	281953	36.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.40%
21) 2-Butanone-d5	3.894	46	343705	103.29	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.29%
24) Chloroform-d	4.351	84	481469	50.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.38%
26) 1,2-Dichloroethane-d4	5.036	65	329890	53.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.42%
32) Benzene-d6	5.055	84	942651	49.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.06%
36) 1,2-Dichloropropane-d6	6.074	67	296155	50.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.00%
41) Toluene-d8	7.319	98	865342	48.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.68%
43) trans-1,3-Dichloroprop...	7.624	79	122842	42.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.62%
47) 2-Hexanone-d5	8.090	63	230602	88.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.12%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	383575	47.69	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.38%
64) 1,2-Dichlorobenzene-d4	11.630	152	382018	52.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.76%
Target Compounds						
					Qvalue	
13) Acetone	2.187	43	15360	8.01	ug/L	95
48) 2-Hexanone	8.151	43	9667	1.80	ug/L	# 62


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020221\
Data File : VV020237.D
Acq On : 02 Feb 2021 14:57
Operator : SY/MD
Sample : M1258-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4Z4

Quant Time: Feb 03 02:29:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 03 02:27:31 2021
Response via : Initial Calibration









#13
Acetone
Concen: 8.01 ug/L
RT: 2.187 min Scan# 341
Delta R.T. 0.007 min
Lab File: VV020237.D
Acq: 02 Feb 2021 14:57

Instrument :
MSVOA_V
ClientSampled :
BG4Z4

Tgt Ion: 43 Resp: 15360
Ion Ratio Lower Upper
43 100
58 32.9 0.0 60.2



#48
2-Hexanone
Concen: 1.80 ug/L
RT: 8.151 min Scan# 2196
Delta R.T. 0.013 min
Lab File: VV020237.D
Acq: 02 Feb 2021 14:57

Tgt Ion: 43 Resp: 9667
Ion Ratio Lower Upper
43 100
58 22.1 41.8 62.6#
57 0.0 15.0 22.4#
100 6.5 8.9 13.3#

