

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020221\
 Data File : VW020251.D
 Acq On : 02 Feb 2021 20:26
 Operator : SY/MD
 Sample : VIBLK82
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK82

Quant Time: Feb 03 02:32:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 03 02:27:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	740053	50.00	ug/L	# 0.00
28) Chlorobenzene-d5	8.855	117	733625	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.254	152	393463	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	166795	32.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.90%
7) Chloroethane-d5	1.569	69	133591	33.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	67.58%#
11) 1,1-Dichloroethene-d2	2.110	63	219171	29.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.54%#
21) 2-Butanone-d5	3.894	46	249711	76.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.98%
24) Chloroform-d	4.354	84	376102	40.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.24%
26) 1,2-Dichloroethane-d4	5.036	65	241390	40.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.64%
32) Benzene-d6	5.055	84	751277	39.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.14%
36) 1,2-Dichloropropane-d6	6.074	67	229491	39.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.46%
41) Toluene-d8	7.318	98	726879	41.12	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.24%
43) trans-1,3-Dichloroprop...	7.624	79	111623	38.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.88%
47) 2-Hexanone-d5	8.090	63	203334	78.68	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.68%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	349141	43.95	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.90%
64) 1,2-Dichlorobenzene-d4	11.630	152	357136	47.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.84%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020221\
Data File : VV020251.D
Acq On : 02 Feb 2021 20:26
Operator : SY/MD
Sample : VIBLK82
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK82

Quant Time: Feb 03 02:32:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 03 02:27:31 2021
Response via : Initial Calibration

