

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020222\
 Data File : VW024519.D
 Acq On : 02 Feb 2022 13:00
 Operator : SY/MD
 Sample : VW0202MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK306

Quant Time: Feb 03 06:24:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM020222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 03 05:56:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	376748	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	338996	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	152188	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	126602	47.322	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.640%
7) Chloroethane-d5	1.568	69	109868	47.116	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.240%
11) 1,1-Dichloroethene-d2	2.108	63	194715	36.531	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.060%
21) 2-Butanone-d5	3.883	46	168302	95.113	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.110%
24) Chloroform-d	4.342	84	229348	46.678	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.360%
26) 1,2-Dichloroethane-d4	5.027	65	151860	48.423	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.840%
32) Benzene-d6	5.047	84	469387	48.697	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.400%
36) 1,2-Dichloropropane-d6	6.066	67	149653	48.772	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.540%
41) Toluene-d8	7.310	98	432268	48.729	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.460%
43) trans-1,3-Dichloroprop...	7.619	79	76366	47.295	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.580%
47) 2-Hexanone-d5	8.085	63	126655	97.184	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.180%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	175130	47.607	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.220%
66) 1,2-Dichlorobenzene-d4	11.622	152	141639	49.191	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020222\
Data File : VV024519.D
Acq On : 02 Feb 2022 13:00
Operator : SY/MD
Sample : VV0202MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK306

Quant Time: Feb 03 06:24:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM020222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 03 05:56:12 2022
Response via : Initial Calibration

