

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020222\
 Data File : VW024520.D
 Acq On : 02 Feb 2022 13:26
 Operator : SY/MD
 Sample : VW0202WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK307

Quant Time: Feb 03 06:24:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM020222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 03 05:56:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	366038	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	335637	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	149394	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	123120	47.367	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	94.740%	
7) Chloroethane-d5	1.568	69	109724	48.431	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.860%	
11) 1,1-Dichloroethene-d2	2.108	63	189455	36.585	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	73.160%	
21) 2-Butanone-d5	3.886	46	165330	96.167	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	96.170%	
24) Chloroform-d	4.346	84	221802	46.463	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.920%	
26) 1,2-Dichloroethane-d4	5.031	65	147965	48.561	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.120%	
32) Benzene-d6	5.047	84	456905	47.877	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.760%	
36) 1,2-Dichloropropane-d6	6.066	67	146320	48.163	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	96.320%	
41) Toluene-d8	7.314	98	418397	47.637	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.280%	
43) trans-1,3-Dichloroprop...	7.619	79	76371	47.771	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	95.540%	
47) 2-Hexanone-d5	8.085	63	119591	92.682	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	92.680%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	169668	46.584	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	93.160%	
66) 1,2-Dichlorobenzene-d4	11.622	152	137253	48.559	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020222\
Data File : VV024520.D
Acq On : 02 Feb 2022 13:26
Operator : SY/MD
Sample : VV0202WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK307

Quant Time: Feb 03 06:24:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM020222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 03 05:56:12 2022
Response via : Initial Calibration

