

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020222\
 Data File : VW024521.D
 Acq On : 02 Feb 2022 13:50
 Operator : SY/MD
 Sample : VW0202WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK308

Quant Time: Feb 03 06:24:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM020222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 03 05:56:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	372923	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	337557	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	145940	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	125214	47.283	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.560%
7) Chloroethane-d5	1.568	69	107138	46.417	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.840%
11) 1,1-Dichloroethene-d2	2.108	63	201023	38.102	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.200%
21) 2-Butanone-d5	3.905	46	142471	81.341	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.340%
24) Chloroform-d	4.349	84	225367	46.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.680%
26) 1,2-Dichloroethane-d4	5.031	65	147537	47.527	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.060%
32) Benzene-d6	5.050	84	462346	48.171	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.340%
36) 1,2-Dichloropropane-d6	6.069	67	147296	48.208	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.420%
41) Toluene-d8	7.313	98	419932	47.540	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.080%
43) trans-1,3-Dichloroprop...	7.619	79	73885	45.953	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.900%
47) 2-Hexanone-d5	8.088	63	125957	97.060	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.060%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	171359	46.781	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.560%
66) 1,2-Dichlorobenzene-d4	11.622	152	135479	49.066	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020222\
Data File : VV024521.D
Acq On : 02 Feb 2022 13:50
Operator : SY/MD
Sample : VV0202WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK308

Quant Time: Feb 03 06:24:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM020222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 03 05:56:12 2022
Response via : Initial Calibration

