

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
 Data File : VV038578.D
 Acq On : 03 Feb 2025 11:31
 Operator : SY/MD
 Sample : VV0203WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK221

Quant Time: Feb 04 07:31:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5233	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5283	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2400	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1983	0.376	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	76.000%
4) 1,2-Dichloroethane-d4	4.969	65	1835	0.436	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	88.000%
7) 1,2-Dichloropropane-d6	5.995	67	2268	0.437	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	88.000%
8) Toluene-d8	7.254	98	4282	0.411	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	82.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1901	0.429	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	86.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
Data File : VV038578.D
Acq On : 03 Feb 2025 11:31
Operator : SY/MD
Sample : VV0203WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK221

Quant Time: Feb 04 07:31:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

