

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
 Data File : VV038579.D
 Acq On : 03 Feb 2025 12:30
 Operator : SY/MD
 Sample : Q1226-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCZK2

Quant Time: Feb 04 07:32:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5324	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5482	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2491	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1937	0.361	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	72.000%		
4) 1,2-Dichloroethane-d4	4.969	65	1996	0.466	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	94.000%		
7) 1,2-Dichloropropane-d6	5.995	67	2387	0.443	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	88.000%		
8) Toluene-d8	7.253	98	4311	0.399	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	80.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2112	0.460	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	92.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
Data File : VV038579.D
Acq On : 03 Feb 2025 12:30
Operator : SY/MD
Sample : Q1226-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCZK2

Quant Time: Feb 04 07:32:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

