

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
 Data File : VV038583.D
 Acq On : 03 Feb 2025 14:04
 Operator : SY/MD
 Sample : Q1226-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCZP1

Quant Time: Feb 04 07:32:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5551	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	5565	0.500	ug/L	#-0.02
11) 1,4-Dichlorobenzene-d4	11.174	152	3338	0.500	ug/L	#-0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	2026	0.362	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	72.000%
4) 1,2-Dichloroethane-d4	4.969	65	1866	0.418	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	84.000%
7) 1,2-Dichloropropane-d6	5.995	67	2043	0.373	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	74.000%
8) Toluene-d8	7.253	98	4183	0.382	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	76.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1978	0.424	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	84.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	4612	0.639	ug/L	96
6) Trichloroethene	5.828	95	730802	158.979	ug/L	91
13) 1,2-Dibromo-3-chloropr...	12.408	75	161	0.268	ug/L #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
Data File : VV038583.D
Acq On : 03 Feb 2025 14:04
Operator : SY/MD
Sample : Q1226-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCZP1

Quant Time: Feb 04 07:32:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

