

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
 Data File : VV038592.D
 Acq On : 03 Feb 2025 17:42
 Operator : SY/MD
 Sample : Q1226-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCZL6

Quant Time: Feb 04 07:34:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6067	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6373	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.174	152	5054	0.500	ug/L	#-0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	2251	0.368	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	74.000%		
4) 1,2-Dichloroethane-d4	4.969	65	2043	0.418	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	84.000%		
7) 1,2-Dichloropropane-d6	5.995	67	2606	0.416	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	84.000%		
8) Toluene-d8	7.253	98	4949	0.394	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	78.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2474	0.463	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	92.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.273	62	264	0.033	ug/L #	49
6) Trichloroethene	5.852	95	963	0.183	ug/L	98
9) 1,2-Dibromoethane	8.305	107	197	0.053	ug/L #	62
13) 1,2-Dibromo-3-chloropr...	12.374	75	1455	1.599	ug/L #	41

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
Data File : VV038592.D
Acq On : 03 Feb 2025 17:42
Operator : SY/MD
Sample : Q1226-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCZL6

Quant Time: Feb 04 07:34:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

