

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
 Data File : VV038596.D
 Acq On : 03 Feb 2025 19:19
 Operator : SY/MD
 Sample : Q1226-25
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCZN7

Quant Time: Feb 04 07:35:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7433	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	7327	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3402	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	2919	0.389	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	78.000%		
4) 1,2-Dichloroethane-d4	4.969	65	2510	0.419	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	84.000%		
7) 1,2-Dichloropropane-d6	5.995	67	3064	0.425	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	86.000%		
8) Toluene-d8	7.253	98	5888	0.408	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	82.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2729	0.444	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	88.000%		
Target Compounds						
6) Trichloroethene	5.876	95	143	0.024	ug/L	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
Data File : VV038596.D
Acq On : 03 Feb 2025 19:19
Operator : SY/MD
Sample : Q1226-25
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCZN7

Quant Time: Feb 04 07:35:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

