

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VW020325\
 Data File : VW038598.D
 Acq On : 03 Feb 2025 20:08
 Operator : SY/MD
 Sample : Q1226-27
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCZP9

Quant Time: Feb 04 07:35:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	4053	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	4295	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2012	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	2014	0.493	ug/L	0.00
Spiked Amount	0.500	Range 40 - 130	Recovery	=	98.000%	
4) 1,2-Dichloroethane-d4	4.969	65	1687	0.517	ug/L	0.00
Spiked Amount	0.500	Range 70 - 130	Recovery	=	104.000%	
7) 1,2-Dichloropropane-d6	5.995	67	1865	0.442	ug/L	0.00
Spiked Amount	0.500	Range 60 - 140	Recovery	=	88.000%	
8) Toluene-d8	7.253	98	3321	0.392	ug/L	0.00
Spiked Amount	0.500	Range 70 - 130	Recovery	=	78.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2191	0.609	ug/L	0.00
Spiked Amount	0.500	Range 65 - 120	Recovery	=	122.000%#	
Target Compounds						
6) Trichloroethene	5.876	95	119	0.034	ug/L #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020325\
Data File : VV038598.D
Acq On : 03 Feb 2025 20:08
Operator : SY/MD
Sample : Q1226-27
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCZP9

Quant Time: Feb 04 07:35:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

