

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020525\
 Data File : VV038627.D
 Acq On : 05 Feb 2025 15:10
 Operator : SY/MD
 Sample : Q1226-22DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 DCZL9DL

Quant Time: Feb 06 07:33:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R. T.	QI on	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	4119	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	4297	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.191	152	2045	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1357	0.327	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	66.000%
4) 1,2-Dichloroethane-d4	4.969	65	1500	0.452	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.000%
7) 1,2-Dichloropropane-d6	5.995	67	1728	0.409	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	82.000%
8) Toluene-d8	7.253	98	2994	0.354	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	70.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1793	0.498	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	100.000%
Target Compounds						
6) Trichloroethene	5.852	95	2102	0.592	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV020525\
 Data File : VV038627.D
 Acq On : 05 Feb 2025 15:10
 Operator : SY/MD
 Sample : Q1226-22DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 DCZL9DL

Quant Time: Feb 06 07:33:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSI M012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

