

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020824\
 Data File : VW034071.D
 Acq On : 08 Feb 2024 20:31
 Operator : SY/MD
 Sample : VW0208WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK211

Quant Time: Feb 09 03:49:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 09 03:48:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	89210	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	83177	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	41728	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	34718	5.025	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 100.400%		
7) Chloroethane-d5	1.532	69	27653	4.569	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 91.400%		
11) 1,1-Dichloroethene-d2	2.060	65	15912	4.776	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 95.600%		
20) 2-Butanone-d5	3.825	46	51663	44.523	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 89.040%		
24) Chloroform-d	4.253	84	56982	4.539	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 90.800%		
26) 1,2-Dichloroethane-d4	4.947	65	29563	5.039	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 100.800%		
32) Benzene-d6	4.963	84	115905	4.645	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 92.800%		
36) 1,2-Dichloropropane-d6	5.992	67	31925	4.577	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 91.600%		
41) Toluene-d8	7.246	98	105665	4.607	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 92.200%		
43) trans-1,3-Dichloroprop...	7.561	79	12262	4.655	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 93.200%		
46) 2-Hexanone-d5	8.030	63	48736	51.831	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 103.660%		
56) 1,1,2,2-Tetrachloroeth...	10.156	84	20885	4.939	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 98.800%		
66) 1,2-Dichlorobenzene-d4	11.564	152	36215	4.772	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 95.400%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020824\
Data File : VV034071.D
Acq On : 08 Feb 2024 20:31
Operator : SY/MD
Sample : VV0208WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK211

Quant Time: Feb 09 03:49:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 09 03:48:51 2024
Response via : Initial Calibration

