

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020824\
 Data File : VW034082.D
 Acq On : 09 Feb 2024 00:57
 Operator : SY/MD
 Sample : P1401-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AJ1

Quant Time: Feb 09 04:48:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 09 03:48:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	82388	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	77112	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	39645	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	20643	3.235	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.800%
7) Chloroethane-d5	1.532	69	20859	3.732	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.600%
11) 1,1-Dichloroethene-d2	2.060	65	11261	3.660	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.200%
20) 2-Butanone-d5	3.815	46	54133	50.515	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.020%
24) Chloroform-d	4.252	84	49208	4.244	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
26) 1,2-Dichloroethane-d4	4.947	65	24954	4.605	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.200%
32) Benzene-d6	4.963	84	92061	3.979	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.600%
36) 1,2-Dichloropropane-d6	5.992	67	27253	4.214	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.200%
41) Toluene-d8	7.246	98	84339	3.966	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	7.561	79	9782	4.006	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.027	63	45320	51.988	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.980%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	19127	4.879	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	32199	4.466	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020824\
Data File : VV034082.D
Acq On : 09 Feb 2024 00:57
Operator : SY/MD
Sample : P1401-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COAJ1

Quant Time: Feb 09 04:48:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 09 03:48:51 2024
Response via : Initial Calibration

