

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020824\
 Data File : VW034083.D
 Acq On : 09 Feb 2024 01:21
 Operator : SY/MD
 Sample : P1401-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AJ2

Quant Time: Feb 09 04:48:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 09 03:48:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	85992	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	80832	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	41280	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	20675	3.104	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.000%
7) Chloroethane-d5	1.532	69	20664	3.542	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.800%
11) 1,1-Dichloroethene-d2	2.060	65	11255	3.504	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.000%
20) 2-Butanone-d5	3.802	46	53417	47.757	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.520%
24) Chloroform-d	4.253	84	50631	4.184	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
26) 1,2-Dichloroethane-d4	4.947	65	25915	4.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.600%
32) Benzene-d6	4.963	84	92590	3.818	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.400%
36) 1,2-Dichloropropane-d6	5.989	67	27556	4.065	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.200%
41) Toluene-d8	7.246	98	85340	3.829	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.600%
43) trans-1,3-Dichloroprop...	7.561	79	9521	3.719	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.400%
46) 2-Hexanone-d5	8.027	63	44206	48.377	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.760%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	18825	4.581	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	32343	4.308	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020824\
Data File : VV034083.D
Acq On : 09 Feb 2024 01:21
Operator : SY/MD
Sample : P1401-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AJ2

Quant Time: Feb 09 04:48:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 09 03:48:51 2024
Response via : Initial Calibration

