


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020824\
Data File : VV034044.D
Acq On : 08 Feb 2024 08:31
Operator : SY/MD
Sample : BFB248
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB248

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Feb 09 03:48:51 2024



AutoFind: Scans 2789, 2790, 2791; Background Corrected with Scan 2777

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	12582	PASS
75	95	30	80	57.9	33701	PASS
95	95	100	100	100.0	58200	PASS
96	95	5	9	6.9	4011	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	85.3	49619	PASS
175	174	5	9	8.0	3986	PASS
176	174	95	105	95.8	47512	PASS
177	176	5	10	6.6	3121	PASS

m/z	Abundance
39.70	132.0

Instrument :
MSVOA_V
ClientSampleId :
BFB248

m/z	Abundance
35.90	724.0
36.90	3094.0
37.90	2715.0
39.00	1181.0
39.80	103.0
43.90	467.0
44.90	624.0
46.90	697.0
47.80	409.0
48.90	2664.0
49.90	13011.0
50.90	3905.0
51.90	183.0
55.00	195.0
55.90	1228.0
56.90	2083.0
59.90	696.0
60.90	3406.0
62.00	3395.0
62.90	2873.0
63.90	200.0
64.90	125.0
66.90	178.0
67.90	6942.0
68.90	7050.0
69.90	564.0
71.90	467.0
73.00	3101.0
74.00	11671.0
75.00	34368.0
76.00	2808.0
76.90	210.0
78.00	284.0
78.90	2725.0
79.90	747.0
80.80	2950.0
81.80	689.0
86.90	2067.0
87.90	2086.0
90.80	294.0
91.90	2313.0
92.90	2707.0
94.00	8037.0
95.00	58472.0
95.90	3996.0
96.90	150.0
103.80	404.0
104.90	106.0
105.70	359.0
110.90	120.0
114.80	140.0
115.70	300.0
116.80	438.0
117.80	339.0
118.80	452.0
127.80	247.0
128.80	122.0
129.70	268.0
130.90	101.0
134.80	105.0
136.70	128.0
140.90	816.0
142.90	921.0
145.80	107.0
147.90	161.0
154.90	172.0
156.90	101.0
173.90	48896.0
174.90	3889.0
175.90	46176.0
176.90	3034.0
177.80	122.0

Instrument :
MSVOA_V
ClientSampleId :
BFB248

m/z	Abundance
35.90	722.0
36.90	3071.0
37.90	2750.0
39.00	1153.0
39.90	106.0
43.90	502.0
44.90	624.0
46.90	698.0
47.90	416.0
48.90	2814.0
49.90	12969.0
50.90	3898.0
51.90	167.0
54.90	215.0
55.90	1276.0
56.90	2068.0
57.90	132.0
59.90	716.0
60.90	3404.0
62.00	3346.0
62.90	2715.0
63.90	214.0
64.90	104.0
66.90	205.0
67.90	7073.0
68.90	7285.0
69.90	584.0
71.90	526.0
73.00	3105.0
73.90	11900.0
75.00	34752.0
76.00	2860.0
76.80	252.0
78.00	286.0
78.80	2735.0
79.90	774.0
80.80	2972.0
81.80	660.0
86.10	109.0
86.90	2003.0
87.90	2093.0
90.80	334.0
91.90	2249.0
92.90	2781.0
94.00	8268.0
95.00	60064.0
96.00	4124.0
96.90	134.0
103.80	397.0
105.70	369.0
110.80	110.0
114.80	125.0
115.80	342.0
116.80	464.0
117.80	319.0
118.80	468.0
127.90	239.0
128.80	104.0
129.80	303.0
134.80	118.0
136.70	115.0
137.00	111.0
140.90	830.0
141.70	109.0
142.90	926.0
145.80	123.0
147.90	167.0
154.90	209.0
156.90	112.0
173.90	51344.0
174.90	4105.0
175.90	48984.0
176.90	3282.0
177.80	114.0

Instrument :
MSVOA_V
ClientSampleId :
BFB248

m/z	Abundance
35.90	634.0
36.90	2846.0
37.90	2495.0
38.90	991.0
39.90	133.0
43.90	495.0
44.90	546.0
46.90	601.0
47.90	433.0
48.90	2658.0
50.00	11765.0
50.90	3596.0
51.90	147.0
54.90	223.0
55.90	1216.0
57.00	1828.0
57.90	126.0
59.90	726.0
60.90	3141.0
61.90	3085.0
62.90	2350.0
63.90	234.0
66.80	224.0
67.90	6525.0
68.90	6692.0
69.90	513.0
71.90	481.0
72.90	2926.0
73.90	11103.0
75.00	31984.0
76.00	2615.0
76.80	290.0
78.10	224.0
78.80	2575.0
79.90	824.0
80.80	2725.0
81.70	555.0
86.80	1849.0
87.90	1989.0
90.80	324.0
91.90	2005.0
92.90	2683.0
94.00	7556.0
95.00	56064.0
96.00	3913.0
96.90	116.0
103.80	351.0
105.70	351.0
106.80	112.0
110.70	107.0
114.80	101.0
115.80	361.0
116.80	507.0
117.80	303.0
118.80	429.0
127.90	227.0
129.80	285.0
134.80	124.0
137.00	112.0
140.80	817.0
142.80	855.0
144.90	105.0
145.80	112.0
147.90	153.0
154.80	215.0
156.80	113.0
171.50	145.0
173.90	48616.0
174.90	3964.0
175.90	47376.0
176.90	3047.0

Instrument :
MSVOA_V
ClientSampleId :
BFB248