

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021122\
 Data File : VW024606.D
 Acq On : 11 Feb 2022 16:11
 Operator : SY/MD
 Sample : N1475-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4M71

Quant Time: Feb 12 03:44:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 03:41:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	98041	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	100182	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49702	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	30255	3.681	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.600%
7) Chloroethane-d5	1.571	69	29607	4.261	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.200%
11) 1,1-Dichloroethene-d2	2.111	63	48274	2.685	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.800%#
20) 2-Butanone-d5	3.896	46	84718	62.098	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.200%
24) Chloroform-d	4.352	84	65178	4.506	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.034	65	35339	5.385	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.053	84	131192	4.615	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.072	67	42052	4.757	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.317	98	114305	4.312	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	7.625	79	16382	4.588	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.800%
46) 2-Hexanone-d5	8.088	63	67767	53.876	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.760%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28370	5.390	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	38639	4.823	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
Data File : VV024606.D
Acq On : 11 Feb 2022 16:11
Operator : SY/MD
Sample : N1475-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4M71

Quant Time: Feb 12 03:44:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 12 03:41:27 2022
Response via : Initial Calibration

