

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
 Data File : VV024610.D
 Acq On : 11 Feb 2022 17:45
 Operator : SY/MD
 Sample : N1479-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBCW2

Quant Time: Feb 12 03:45:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 03:41:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	123265	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	123000	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60969	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	43772	4.236	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.800%
7) Chloroethane-d5	1.568	69	41541	4.756	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.108	63	70564	3.122	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.400%
20) 2-Butanone-d5	3.892	46	92918	54.171	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.340%
24) Chloroform-d	4.349	84	84465	4.645	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.034	65	41381	5.015	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	5.050	84	169472	4.856	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	6.069	67	53275	4.908	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.317	98	155403	4.774	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
43) trans-1,3-Dichloroprop...	7.625	79	18774	4.283	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.600%
46) 2-Hexanone-d5	8.088	63	76050	49.245	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.480%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31335	4.849	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	49221	5.008	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.200%
Target Compounds					Qvalue	
25) Chloroform	4.372	83	1629	0.072	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
Data File : VV024610.D
Acq On : 11 Feb 2022 17:45
Operator : SY/MD
Sample : N1479-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBCW2

Quant Time: Feb 12 03:45:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 12 03:41:27 2022
Response via : Initial Calibration

