

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021122\
 Data File : VW024621.D
 Acq On : 11 Feb 2022 22:58
 Operator : SY/MD
 Sample : N1512-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB3

Quant Time: Feb 12 07:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 07:08:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	118672	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	120632	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60053	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	39692	3.990	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	79.800%	
7) Chloroethane-d5	1.571	69	37137	4.416	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.400%	
11) 1,1-Dichloroethene-d2	2.111	63	62836	2.888	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.800%#	
20) 2-Butanone-d5	3.902	46	89341	54.102	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.200%	
24) Chloroform-d	4.352	84	80030	4.571	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.400%	
26) 1,2-Dichloroethane-d4	5.037	65	40753	5.130	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.600%	
32) Benzene-d6	5.053	84	163812	4.786	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.069	67	50988	4.790	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.800%	
41) Toluene-d8	7.316	98	144794	4.536	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.800%	
43) trans-1,3-Dichloroprop...	7.625	79	17859	4.154	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.000%	
46) 2-Hexanone-d5	8.088	63	71812	47.413	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.820%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	30682	4.841	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	47370	4.893	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.800%	
Target Compounds						
						Qvalue
9) Trichlorofluoromethane	1.757	101	1392	0.062	ug/L	97
25) Chloroform	4.371	83	2355	0.108	ug/L	99
47) Tetrachloroethene	7.976	164	34034	3.957	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
Data File : VV024621.D
Acq On : 11 Feb 2022 22:58
Operator : SY/MD
Sample : N1512-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB3

Quant Time: Feb 12 07:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 12 07:08:39 2022
Response via : Initial Calibration

