

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021122\
 Data File : VW024627.D
 Acq On : 12 Feb 2022 01:18
 Operator : SY/MD
 Sample : N1512-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA7

Quant Time: Feb 12 07:10:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 07:08:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	114087	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	115280	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56122	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	35222	3.683	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	73.600%	
7) Chloroethane-d5	1.571	69	34846	4.310	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	86.200%	
11) 1,1-Dichloroethene-d2	2.111	63	58555	2.799	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	56.000%#	
20) 2-Butanone-d5	3.902	46	88100	55.494	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	110.980%	
24) Chloroform-d	4.349	84	76184	4.526	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.600%	
26) 1,2-Dichloroethane-d4	5.037	65	38492	5.040	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.800%	
32) Benzene-d6	5.053	84	153570	4.695	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
36) 1,2-Dichloropropane-d6	6.069	67	48549	4.772	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.400%	
41) Toluene-d8	7.317	98	137111	4.494	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	89.800%	
43) trans-1,3-Dichloroprop...	7.625	79	17014	4.141	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.800%	
46) 2-Hexanone-d5	8.088	63	71700	49.537	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.080%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	29867	4.931	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	43922	4.855	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.000%	
Target Compounds						
17) Methyl tert-butyl Ether	2.770	73	23254	0.955	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
 Data File : VV024627.D
 Acq On : 12 Feb 2022 01:18
 Operator : SY/MD
 Sample : N1512-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 C0AA7

Quant Time: Feb 12 07:10:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 07:08:39 2022
 Response via : Initial Calibration

