

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021122\
 Data File : VW024636.D
 Acq On : 12 Feb 2022 04:48
 Operator : SY/MD
 Sample : N1480-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBCX3

Quant Time: Feb 12 07:11:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 07:08:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	113834	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	115078	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	57417	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	35367	3.706	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	74.200%	
7) Chloroethane-d5	1.571	69	34049	4.221	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.400%	
11) 1,1-Dichloroethene-d2	2.111	63	56946	2.728	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	54.600%#	
20) 2-Butanone-d5	3.895	46	86317	54.492	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.980%	
24) Chloroform-d	4.352	84	76468	4.553	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.000%	
26) 1,2-Dichloroethane-d4	5.034	65	37013	4.857	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.053	84	150358	4.605	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.000%	
36) 1,2-Dichloropropane-d6	6.072	67	46740	4.603	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.000%	
41) Toluene-d8	7.316	98	133590	4.387	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.800%	
43) trans-1,3-Dichloroprop...	7.625	79	16989	4.143	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.800%	
46) 2-Hexanone-d5	8.088	63	69585	48.160	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.320%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	30200	4.995	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	44051	4.759	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
Data File : VV024636.D
Acq On : 12 Feb 2022 04:48
Operator : SY/MD
Sample : N1480-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBCX3

Quant Time: Feb 12 07:11:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 12 07:08:39 2022
Response via : Initial Calibration

