

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021122\
 Data File : VW024638.D
 Acq On : 12 Feb 2022 05:35
 Operator : SY/MD
 Sample : N1480-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBCX7

Quant Time: Feb 12 07:12:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 12 07:08:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	116268	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	118216	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	58815	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	34582	3.548	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.000%	
7) Chloroethane-d5	1.571	69	33965	4.122	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	82.400%	
11) 1,1-Dichloroethene-d2	2.111	63	56375	2.644	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	52.800%#	
20) 2-Butanone-d5	3.896	46	85069	52.580	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.160%	
24) Chloroform-d	4.352	84	76914	4.484	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.600%	
26) 1,2-Dichloroethane-d4	5.034	65	37613	4.833	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.600%	
32) Benzene-d6	5.053	84	150937	4.500	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.000%	
36) 1,2-Dichloropropane-d6	6.069	67	49120	4.709	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.200%	
41) Toluene-d8	7.317	98	133162	4.257	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.200%	
43) trans-1,3-Dichloroprop...	7.625	79	15789	3.748	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.000%	
46) 2-Hexanone-d5	8.088	63	71046	47.866	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.740%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29746	4.789	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	44486	4.692	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021122\
Data File : VV024638.D
Acq On : 12 Feb 2022 05:35
Operator : SY/MD
Sample : N1480-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBCX7

Quant Time: Feb 12 07:12:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 12 07:08:39 2022
Response via : Initial Calibration

