



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021324\
Data File : VV034143.D
Acq On : 13 Feb 2024 08:49
Operator : SY/MD
Sample : BFB251
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB251

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Wed Feb 14 01:41:00 2024



AutoFind: Scans 2789, 2790, 2791; Background Corrected with Scan 2777

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	12217	PASS
75	95	30	80	57.4	32584	PASS
95	95	100	100	100.0	56773	PASS
96	95	5	9	6.7	3812	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	86.3	49000	PASS
175	174	5	9	7.8	3832	PASS
176	174	95	105	99.0	48528	PASS
177	176	5	10	6.3	3065	PASS

m/z	Abundance
39.80	134.0
43.80	132.0

Instrument :
MSVOA_V
ClientSampleId :
BFB251

m/z	Abundance
36.00	546.0
36.90	3381.0
37.90	2726.0
38.90	973.0
43.90	501.0
44.90	567.0
46.90	671.0
47.80	432.0
48.90	2875.0
49.90	12417.0
50.90	3802.0
55.00	195.0
55.90	1182.0
56.90	1939.0
60.00	639.0
60.90	3377.0
61.90	3351.0
62.90	2550.0
64.00	305.0
64.90	142.0
66.90	256.0
67.90	6615.0
69.00	7092.0
69.90	445.0
71.80	376.0
72.90	3407.0
73.90	12311.0
75.00	32968.0
75.90	3058.0
76.90	414.0
78.00	208.0
78.90	2772.0
79.90	790.0
80.90	3196.0
81.90	700.0
85.90	132.0
86.90	2384.0
87.90	2114.0
88.60	104.0
90.80	303.0
91.90	2162.0
92.90	2964.0
93.90	7807.0
95.00	57816.0
96.00	3962.0
96.90	137.0
103.70	469.0
104.70	101.0
105.80	441.0
115.80	297.0
116.70	515.0
117.90	335.0
118.80	458.0
127.90	302.0
129.80	254.0
134.90	142.0
136.80	158.0
140.80	729.0
142.80	1003.0
144.90	108.0
147.80	158.0
154.80	179.0
156.90	118.0
172.20	270.0
173.90	48536.0
174.90	3873.0
175.90	48304.0
176.90	3192.0

Instrument :
MSVOA_V
ClientSampleId :
BFB251

m/z	Abundance
35.90	544.0
36.90	3425.0
37.90	2647.0
38.90	934.0
43.90	456.0
44.90	609.0
46.90	620.0
47.80	409.0
49.00	2894.0
49.90	12644.0
50.90	3956.0
54.90	222.0
55.90	1107.0
56.90	1921.0
59.90	622.0
60.90	3360.0
61.90	3377.0
62.90	2655.0
64.10	247.0
65.00	157.0
66.90	217.0
67.90	6673.0
68.90	7250.0
69.90	441.0
71.90	425.0
72.90	3323.0
74.00	12393.0
74.90	33664.0
76.00	3069.0
76.80	400.0
78.00	213.0
78.90	2752.0
79.80	802.0
80.90	3120.0
81.90	740.0
85.90	146.0
86.90	2498.0
87.90	1980.0
90.80	338.0
91.90	2264.0
92.90	2935.0
93.90	7904.0
95.00	58640.0
96.00	3954.0
96.90	121.0
103.70	481.0
105.80	416.0
115.80	355.0
116.80	510.0
117.90	357.0
118.80	520.0
127.90	299.0
128.70	139.0
129.80	254.0
134.90	150.0
136.70	165.0
140.80	849.0
142.80	991.0
144.80	117.0
147.80	179.0
154.80	188.0
156.80	142.0
172.00	240.0
173.90	50616.0
174.90	3935.0
175.90	50224.0
176.90	3158.0

Instrument :
MSVOA_V
ClientSampleId :
BFB251

m/z	Abundance
35.90	513.0
36.90	3134.0
37.90	2418.0
39.00	851.0
43.80	436.0
44.90	577.0
46.90	485.0
47.90	372.0
49.00	2762.0
49.90	11589.0
50.90	3775.0
52.00	104.0
54.90	225.0
55.90	1017.0
56.90	1741.0
59.90	549.0
60.90	3070.0
61.90	3230.0
62.90	2611.0
63.80	193.0
64.10	192.0
65.00	138.0
66.80	157.0
67.90	6179.0
68.90	6564.0
69.90	394.0
71.90	427.0
72.90	3029.0
74.00	11306.0
75.00	31120.0
76.00	2814.0
76.90	367.0
78.00	219.0
78.90	2507.0
79.80	735.0
80.90	2793.0
81.90	716.0
86.00	133.0
86.90	2148.0
87.90	1817.0
90.80	337.0
91.90	2044.0
92.90	2723.0
93.90	7299.0
95.00	53864.0
96.00	3520.0
103.80	428.0
105.80	367.0
115.80	364.0
116.80	478.0
117.90	318.0
118.80	515.0
127.80	265.0
128.70	148.0
129.80	247.0
134.80	144.0
136.70	150.0
140.80	879.0
142.80	919.0
144.80	112.0
147.80	175.0
150.00	100.0
154.80	170.0
156.80	164.0
171.80	249.0
173.90	47848.0
174.90	3687.0
175.90	47056.0
176.90	2844.0

Instrument :
MSVOA_V
ClientSampleId :
BFB251