

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021422\
 Data File : VW024649.D
 Acq On : 14 Feb 2022 14:30
 Operator : SY/MD
 Sample : N1476-10DL 4X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFY00DL

Quant Time: Feb 15 06:27:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 15 06:26:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	116281	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	118617	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	58943	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	48625	4.989	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	99.800%	
7) Chloroethane-d5	1.571	69	45321	5.500	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	110.000%	
11) 1,1-Dichloroethene-d2	2.111	63	78039	3.660	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.200%	
20) 2-Butanone-d5	3.896	46	97680	60.368	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.740%	
24) Chloroform-d	4.349	84	91243	5.319	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	106.400%	
26) 1,2-Dichloroethane-d4	5.034	65	44857	5.763	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.200%	
32) Benzene-d6	5.050	84	194887	5.791	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	115.800%	
36) 1,2-Dichloropropane-d6	6.069	67	58301	5.570	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	111.400%	
41) Toluene-d8	7.317	98	173470	5.526	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.600%	
43) trans-1,3-Dichloroprop...	7.625	79	20737	4.906	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.200%	
46) 2-Hexanone-d5	8.088	63	77393	51.966	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.940%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	32912	5.281	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	54323	5.717	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	114.400%	
Target Compounds						
17) Methyl tert-butyl Ether	2.770	73	2369	0.095	ug/L	90
22) cis-1,2-Dichloroethene	3.915	96	7264	0.609	ug/L	94
25) Chloroform	4.384	83	3364	0.157	ug/L	88
34) Trichloroethene	5.915	95	77137	6.051	ug/L	87
47) Tetrachloroethene	7.976	164	48136	5.692	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021422\
Data File : VV024649.D
Acq On : 14 Feb 2022 14:30
Operator : SY/MD
Sample : N1476-10DL 4X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY00DL

Quant Time: Feb 15 06:27:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 15 06:26:19 2022
Response via : Initial Calibration

