

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021422\
 Data File : VW024660.D
 Acq On : 14 Feb 2022 18:46
 Operator : SY/MD
 Sample : N1475-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4M91

Quant Time: Feb 15 06:30:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 15 06:26:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	111359	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	112986	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56386	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	44959	4.816	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.400%
7) Chloroethane-d5	1.571	69	42297	5.360	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.200%
11) 1,1-Dichloroethene-d2	2.111	63	71817	3.517	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.400%
20) 2-Butanone-d5	3.896	46	89809	57.956	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.920%
24) Chloroform-d	4.349	84	87911	5.351	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
26) 1,2-Dichloroethane-d4	5.037	65	42933	5.759	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.200%
32) Benzene-d6	5.050	84	182284	5.686	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.800%
36) 1,2-Dichloropropane-d6	6.069	67	56290	5.646	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.000%
41) Toluene-d8	7.317	98	163712	5.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.600%
43) trans-1,3-Dichloroprop...	7.625	79	19498	4.842	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.800%
46) 2-Hexanone-d5	8.088	63	74366	52.422	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.840%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	29715	5.005	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	50819	5.591	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.800%
Target Compounds						
					Qvalue	
13) Acetone	2.182	43	4965	4.060	ug/L	82
25) Chloroform	4.375	83	4382	0.214	ug/L	98
42) Toluene	7.384	91	2153865	44.838	ug/L	100
52) Ethylbenzene	9.021	91	6122	0.115	ug/L	98
53) m,p-xylene	9.143	106	7667	0.379	ug/L	97
54) o-xylene	9.548	106	2633	0.135	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021422\
Data File : VV024660.D
Acq On : 14 Feb 2022 18:46
Operator : SY/MD
Sample : N1475-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4M91

Quant Time: Feb 15 06:30:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 15 06:26:19 2022
Response via : Initial Calibration

