

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021424\
 Data File : VW034179.D
 Acq On : 14 Feb 2024 14:15
 Operator : SY/MD
 Sample : P1455-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AG0

Quant Time: Feb 15 01:30:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 15 01:26:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	78502	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	72272	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	37102	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	20338	3.345	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.000%
7) Chloroethane-d5	1.535	69	19552	3.672	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.400%
11) 1,1-Dichloroethene-d2	2.060	65	10562	3.602	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.000%
20) 2-Butanone-d5	3.825	46	39978	39.153	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.300%
24) Chloroform-d	4.252	84	45633	4.131	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.600%
26) 1,2-Dichloroethane-d4	4.947	65	23120	4.478	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	4.963	84	83561	3.854	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.000%
36) 1,2-Dichloropropane-d6	5.992	67	23892	3.942	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.800%
41) Toluene-d8	7.246	98	75997	3.814	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.200%
43) trans-1,3-Dichloroprop...	7.561	79	9594	4.192	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.030	63	31368	38.393	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.780%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	15204	4.138	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.800%
66) 1,2-Dichlorobenzene-d4	11.561	152	28542	4.230	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021424\
Data File : VV034179.D
Acq On : 14 Feb 2024 14:15
Operator : SY/MD
Sample : P1455-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AG0

Quant Time: Feb 15 01:30:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 15 01:26:50 2024
Response via : Initial Calibration

