

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021622\
 Data File : VV024692.D
 Acq On : 16 Feb 2022 10:53
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC005

Quant Time: Feb 16 23:32:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 03:37:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	5.000	5.338	-6.8	110	0.00
3 T	Chloromethane	5.000	5.072	-1.4	107	0.00
4 S	Vinyl Chloride-d3	5.000	3.941	21.2	85	0.00
5 T	Vinyl chloride	5.000	4.980	0.4	106	0.00
6 T	Bromomethane	5.000	5.024	-0.5	106	0.00
7 S	Chloroethane-d5	5.000	4.520	9.6	96	0.00
8 T	Chloroethane	5.000	5.219	-4.4	111	0.00
9 T	Trichlorofluoromethane	5.000	5.309	-6.2	109	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	5.369	-7.4	111	0.00
11 S	1,1-Dichloroethene-d2	5.000	4.333	13.3	91	0.00
12 T	1,1-Dichloroethene	5.000	5.119	-2.4	109	0.00
13 T	Acetone	50.000	52.393	-4.8	134	0.00
14 T	Carbon disulfide	5.000	4.766	4.7	105	0.00
15 T	Methyl Acetate	5.000	4.711	5.8	132	0.00
16 T	Methylene chloride	5.000	4.942	1.2	111	0.00
17 T	Methyl tert-butyl Ether	5.000	5.011	-0.2	106	0.00
18 T	trans-1,2-Dichloroethene	5.000	5.246	-4.9	111	0.00
19 T	1,1-Dichloroethane	5.000	5.493	-9.9	113	0.00
20 S	2-Butanone-d5	50.000	54.662	-9.3	117	0.00
21 T	2-Butanone	50.000	53.990	-8.0	117	0.00
22 T	cis-1,2-Dichloroethene	5.000	5.446	-8.9	111	0.00
23 T	Bromochloromethane	5.000	5.562	-11.2	113	0.00
24 S	Chloroform-d	5.000	5.181	-3.6	106	0.00
25 T	Chloroform	5.000	5.388	-7.8	111	0.00
26 S	1,2-Dichloroethane-d4	5.000	5.027	-0.5	102	0.00
27 T	1,2-Dichloroethane	5.000	5.311	-6.2	109	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	110	0.00
29 T	1,1,1-Trichloroethane	5.000	5.185	-3.7	109	0.00
30 T	Cyclohexane	5.000	5.171	-3.4	109	0.00
31 T	Carbon tetrachloride	5.000	5.281	-5.6	111	0.00
32 S	Benzene-d6	5.000	4.805	3.9	102	0.00
33 T	Benzene	5.000	5.327	-6.5	112	0.00
34 T	Trichloroethene	5.000	5.228	-4.6	109	0.00
35 T	Methylcyclohexane	5.000	5.335	-6.7	110	0.00
36 S	1,2-Dichloropropane-d6	5.000	4.717	5.7	104	0.00
37 T	1,2-Dichloropropane	5.000	5.300	-6.0	114	0.00
38 T	Bromodichloromethane	5.000	5.105	-2.1	107	0.00
39 T	cis-1,3-Dichloropropene	5.000	4.965	0.7	103	0.00
40 T	4-Methyl-2-pentanone	50.000	49.708	0.6	104	0.00
41 S	Toluene-d8	5.000	4.632	7.4	98	0.00
42 T	Toluene	5.000	5.331	-6.6	113	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	4.472	10.6	96	0.00
44 T	trans-1,3-Dichloropropene	5.000	4.902	2.0	102	0.00
45 T	1,1,2-Trichloroethane	5.000	5.286	-5.7	111	0.00
46 S	2-Hexanone-d5	50.000	48.576	2.8	102	0.00
47 T	Tetrachloroethene	5.000	5.431	-8.6	116	0.00
48 T	2-Hexanone	50.000	51.295	-2.6	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021622\
 Data File : VV024692.D
 Acq On : 16 Feb 2022 10:53
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC005

Quant Time: Feb 16 23:32:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 03:37:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	Dibromochloromethane	5.000	5.077	-1.5	108	0.00
50 T	1,2-Dibromoethane	5.000	5.194	-3.9	109	0.00
51 T	Chlorobenzene	5.000	5.382	-7.6	115	0.00
52 T	Ethylbenzene	5.000	5.211	-4.2	110	0.00
53 T	m,p-xylene	5.000	5.260	-5.2	110	0.00
54 T	o-xylene	5.000	5.278	-5.6	112	0.00
55 T	Styrene	5.000	5.331	-6.6	113	0.00
56 S	1,1,2,2-Tetrachloroethane-d	5.000	5.184	-3.7	110	0.00
57 T	1,1,2,2-Tetrachloroethane	5.000	5.376	-7.5	113	0.00
58 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	112	0.00
59 T	Bromoform	5.000	4.972	0.6	108	0.00
60 T	Isopropylbenzene	5.000	5.166	-3.3	111	0.00
61 T	1,2,3-Trichloropropane	5.000	5.105	-2.1	107	0.00
62 T	1,3,5-Trimethylbenzene	5.000	5.188	-3.8	112	0.00
63 T	1,2,4-Trimethylbenzene	5.000	5.199	-4.0	112	0.00
64 T	1,3-Dichlorobenzene	5.000	5.329	-6.6	115	0.00
65 T	1,4-Dichlorobenzene	5.000	5.332	-6.6	115	0.00
66 S	1,2-Dichlorobenzene-d4	5.000	4.678	6.4	102	0.00
67 T	1,2-Dichlorobenzene	5.000	5.328	-6.6	116	0.00
68 T	1,2-Dibromo-3-chloropropane	5.000	4.776	4.5	106	0.00
69	1,3,5-Trichlorobenzene	5.000	5.384	-7.7	115	0.00
70 T	1,2,4-trichlorobenzene	5.000	5.169	-3.4	108	0.00
71	Naphthalene	5.000	4.665	6.7	99	0.00
72 T	1,2,3-Trichlorobenzene	5.000	5.162	-3.2	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0