

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021622\
 Data File : VW024692.D
 Acq On : 16 Feb 2022 10:53
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC005

Quant Time: Feb 16 23:32:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 03:37:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.657	0.701	-6.7	110	0.00
3 T	Chloromethane	0.646	0.656	-1.5	107	0.00
4 S	Vinyl Chloride-d3	0.419	0.330	21.2	85	0.00
5 T	Vinyl chloride	0.696	0.693	0.4	106	0.00
6 T	Bromomethane	0.429	0.431	-0.5	106	0.00
7 S	Chloroethane-d5	0.354	0.320	9.6	96	0.00
8 T	Chloroethane	0.426	0.445	-4.5	111	0.00
9 T	Trichlorofluoromethane	0.941	0.999	-6.2	109	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.525	0.563	-7.2	111	0.00
11 S	1,1-Dichloroethene-d2	0.917	0.795	13.3	91	0.00
12 T	1,1-Dichloroethene	0.518	0.530	-2.3	109	0.00
13 T	Acetone	0.055	0.058	-5.5	134	0.00
14 T	Carbon disulfide	1.645	1.568	4.7	105	0.00
15 T	Methyl Acetate	0.157	0.148	5.7	132	0.00
16 T	Methylene chloride	0.513	0.507	1.2	111	0.00
17 T	Methyl tert-butyl Ether	1.067	1.069	-0.2	106	0.00
18 T	trans-1,2-Dichloroethene	0.509	0.534	-4.9	111	0.00
19 T	1,1-Dichloroethane	0.878	0.965	-9.9	113	0.00
20 S	2-Butanone-d5	0.070	0.076	-8.6	117	0.00
21 T	2-Butanone	0.091	0.098	-7.7	117	0.00
22 T	cis-1,2-Dichloroethene	0.513	0.558	-8.8	111	0.00
23 T	Bromochloromethane	0.195	0.217	-11.3	113	0.00
24 S	Chloroform-d	0.738	0.764	-3.5	106	0.00
25 T	Chloroform	0.919	0.991	-7.8	111	0.00
26 S	1,2-Dichloroethane-d4	0.335	0.337	-0.6	102	0.00
27 T	1,2-Dichloroethane	0.504	0.535	-6.2	109	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
29 T	1,1,1-Trichloroethane	0.820	0.851	-3.8	109	0.00
30 T	Cyclohexane	0.845	0.874	-3.4	109	0.00
31 T	Carbon tetrachloride	0.695	0.734	-5.6	111	0.00
32 S	Benzene-d6	1.419	1.363	3.9	102	0.00
33 T	Benzene	1.951	2.079	-6.6	112	0.00
34 T	Trichloroethene	0.537	0.562	-4.7	109	0.00
35 T	Methylcyclohexane	0.889	0.949	-6.7	110	0.00
36 S	1,2-Dichloropropane-d6	0.441	0.416	5.7	104	0.00
37 T	1,2-Dichloropropane	0.473	0.501	-5.9	114	0.00
38 T	Bromodichloromethane	0.629	0.642	-2.1	107	0.00
39 T	cis-1,3-Dichloropropene	0.721	0.716	0.7	103	0.00
40 T	4-Methyl-2-pentanone	0.244	0.243	0.4	104	0.00
41 S	Toluene-d8	1.323	1.226	7.3	98	0.00
42 T	Toluene	2.126	2.266	-6.6	113	0.00
43 S	trans-1,3-Dichloropropene-d	0.178	0.159	10.7	96	0.00
44 T	trans-1,3-Dichloropropene	0.595	0.584	1.8	102	0.00
45 T	1,1,2-Trichloroethane	0.307	0.324	-5.5	111	0.00
46 S	2-Hexanone-d5	0.063	0.061	3.2	102	0.00
47 T	Tetrachloroethene	0.356	0.387	-8.7	116	0.00
48 T	2-Hexanone	0.167	0.171	-2.4	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021622\
 Data File : VV024692.D
 Acq On : 16 Feb 2022 10:53
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC005

Quant Time: Feb 16 23:32:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 03:37:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Dibromochloromethane	0.370	0.376	-1.6	108	0.00
50 T	1,2-Dibromoethane	0.288	0.299	-3.8	109	0.00
51 T	Chlorobenzene	1.303	1.402	-7.6	115	0.00
52 T	Ethylbenzene	2.365	2.465	-4.2	110	0.00
53 T	m,p-xylene	0.896	0.942	-5.1	110	0.00
54 T	o-xylene	0.864	0.912	-5.6	112	0.00
55 T	Styrene	1.418	1.512	-6.6	113	0.00
56 S	1,1,2,2-Tetrachloroethane-d	0.263	0.272	-3.4	110	0.00
57 T	1,1,2,2-Tetrachloroethane	0.310	0.334	-7.7	113	0.00
58 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
59 T	Bromoform	0.319	0.318	0.3	108	0.00
60 T	Isopropylbenzene	4.535	4.686	-3.3	111	0.00
61 T	1,2,3-Trichloropropane	0.493	0.503	-2.0	107	0.00
62 T	1,3,5-Trimethylbenzene	3.883	4.029	-3.8	112	0.00
63 T	1,2,4-Trimethylbenzene	3.861	4.014	-4.0	112	0.00
64 T	1,3-Dichlorobenzene	1.846	1.967	-6.6	115	0.00
65 T	1,4-Dichlorobenzene	1.833	1.955	-6.7	115	0.00
66 S	1,2-Dichlorobenzene-d4	0.806	0.754	6.5	102	0.00
67 T	1,2-Dichlorobenzene	1.639	1.746	-6.5	116	0.00
68 T	1,2-Dibromo-3-chloropropane	0.096	0.091	5.2	106	0.00
69	1,3,5-Trichlorobenzene	1.313	1.414	-7.7	115	0.00
70 T	1,2,4-trichlorobenzene	1.061	1.096	-3.3	108	0.00
71	Naphthalene	1.921	1.792	6.7	99	0.00
72 T	1,2,3-Trichlorobenzene	0.892	0.921	-3.3	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0