

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021624\
 Data File : VW034213.D
 Acq On : 16 Feb 2024 13:34
 Operator : SY/MD
 Sample : P1501-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E10B3

Quant Time: Feb 19 00:44:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Feb 19 00:41:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	85370	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	79744	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	38102	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	27577	4.171	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.400%
7) Chloroethane-d5	1.532	69	23604	4.076	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.600%
11) 1,1-Dichloroethene-d2	2.056	65	13413	4.207	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.200%
20) 2-Butanone-d5	3.834	46	34418	30.996	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	62.000%
24) Chloroform-d	4.249	84	49466	4.117	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.400%
26) 1,2-Dichloroethane-d4	4.944	65	25084	4.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
32) Benzene-d6	4.960	84	95040	3.972	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
36) 1,2-Dichloropropane-d6	5.992	67	25136	3.758	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.200%
41) Toluene-d8	7.243	98	90005	4.093	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.800%
43) trans-1,3-Dichloroprop...	7.561	79	9803	3.882	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.600%
46) 2-Hexanone-d5	8.030	63	32674	36.245	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.480%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	16002	3.947	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	30582	4.413	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021624\
Data File : VV034213.D
Acq On : 16 Feb 2024 13:34
Operator : SY/MD
Sample : P1501-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E10B3

Quant Time: Feb 19 00:44:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Feb 19 00:41:47 2024
Response via : Initial Calibration

