

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021624\
 Data File : VW034214.D
 Acq On : 16 Feb 2024 13:58
 Operator : SY/MD
 Sample : P1501-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E10B4

Quant Time: Feb 19 00:44:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Feb 19 00:41:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	74448	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	67353	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	33102	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	24747	4.292	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.800%	
7) Chloroethane-d5	1.532	69	21791	4.315	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	86.200%	
11) 1,1-Dichloroethene-d2	2.060	65	12504	4.497	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.000%	
20) 2-Butanone-d5	3.841	46	35496	36.656	ug/L	0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	73.320%	
24) Chloroform-d	4.253	84	45541	4.347	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.000%	
26) 1,2-Dichloroethane-d4	4.947	65	23990	4.900	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
32) Benzene-d6	4.960	84	87769	4.343	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6	5.992	67	23529	4.165	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	83.400%	
41) Toluene-d8	7.243	98	82812	4.459	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	89.200%	
43) trans-1,3-Dichloroprop...	7.561	79	8917	4.181	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.600%	
46) 2-Hexanone-d5	8.034	63	30016	39.422	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	78.840%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	14607	4.265	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.400%	
66) 1,2-Dichlorobenzene-d4	11.561	152	27636	4.591	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021624\
Data File : VV034214.D
Acq On : 16 Feb 2024 13:58
Operator : SY/MD
Sample : P1501-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E10B4

Quant Time: Feb 19 00:44:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Feb 19 00:41:47 2024
Response via : Initial Calibration

