

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021722\
 Data File : VW024725.D
 Acq On : 17 Feb 2022 16:27
 Operator : SY/MD
 Sample : N1480-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 18 07:34:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 18 07:32:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	172665	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	163298	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	71033	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	72438	4.946	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	99.000%	
7) Chloroethane-d5	1.568	69	60399	5.027	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	100.600%	
11) 1,1-Dichloroethene-d2	2.108	63	97247	3.623	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.400%	
20) 2-Butanone-d5	3.896	46	108821	52.020	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.040%	
24) Chloroform-d	4.349	84	108379	4.743	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.800%	
26) 1,2-Dichloroethane-d4	5.034	65	52272	5.132	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.600%	
32) Benzene-d6	5.053	84	231731	4.919	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
36) 1,2-Dichloropropane-d6	6.069	67	68979	4.976	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.600%	
41) Toluene-d8	7.317	98	203812	4.761	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.200%	
43) trans-1,3-Dichloroprop...	7.625	79	21888	4.739	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.800%	
46) 2-Hexanone-d5	8.088	63	86150	47.406	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.820%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	38609	4.621	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	62700	5.101	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.000%	
Target Compounds						Qvalue
25) Chloroform	4.375	83	3816	0.152	ug/L	98
51) Chlorobenzene	8.883	112	4040	0.110	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021722\
Data File : VV024725.D
Acq On : 17 Feb 2022 16:27
Operator : SY/MD
Sample : N1480-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 18 07:34:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 18 07:32:51 2022
Response via : Initial Calibration

