

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022024\
 Data File : VV034275.D
 Acq On : 20 Feb 2024 17:20
 Operator : SY/MD
 Sample : P1498-11 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB2

Quant Time: Feb 21 01:46:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 21 01:41:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	66463	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	61307	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	29344	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	20957	4.071	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.400%
7) Chloroethane-d5	1.532	69	19622	4.352	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.000%
11) 1,1-Dichloroethene-d2	2.060	65	10753	4.332	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.600%
20) 2-Butanone-d5	3.831	46	38786	44.866	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.740%
24) Chloroform-d	4.249	84	46597	4.982	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
26) 1,2-Dichloroethane-d4	4.947	65	22821	5.221	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	4.960	84	84778	4.609	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
36) 1,2-Dichloropropane-d6	5.992	67	24175	4.702	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.243	98	78420	4.639	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
43) trans-1,3-Dichloroprop...	7.561	79	8070	4.157	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.200%
46) 2-Hexanone-d5	8.030	63	34578	49.892	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.780%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	15879	5.094	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	11.561	152	28131	5.271	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.400%
Target Compounds						
					Qvalue	
22) cis-1,2-Dichloroethene	3.818	96	12249	2.283	ug/L	96
34) Trichloroethene	5.831	95	54445	10.231	ug/L	96
47) Tetrachloroethene	7.908	164	3727	0.862	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022024\
Data File : VV034275.D
Acq On : 20 Feb 2024 17:20
Operator : SY/MD
Sample : P1498-11 20X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB2

Quant Time: Feb 21 01:46:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 21 01:41:15 2024
Response via : Initial Calibration

