

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW022024\  
 Data File : VW034278.D  
 Acq On : 20 Feb 2024 18:34  
 Operator : SY/MD  
 Sample : P1498-15 500X  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 24 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_V  
**ClientSampleId :**  
 C0AH3

Quant Time: Feb 21 01:46:48 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR012524WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Feb 21 01:41:15 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	68980	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	62498	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	30233	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	20515	3.840	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.800%	
7) Chloroethane-d5	1.532	69	19607	4.190	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	83.800%	
11) 1,1-Dichloroethene-d2	2.060	65	10592	4.111	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.200%	
20) 2-Butanone-d5	3.831	46	34480	38.429	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	76.860%	
24) Chloroform-d	4.253	84	47017	4.843	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
26) 1,2-Dichloroethane-d4	4.947	65	24389	5.376	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.600%	
32) Benzene-d6	4.963	84	86194	4.597	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.000%	
36) 1,2-Dichloropropane-d6	5.989	67	24494	4.673	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.400%	
41) Toluene-d8	7.246	98	78731	4.569	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.400%	
43) trans-1,3-Dichloroprop...	7.564	79	8521	4.305	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.200%	
46) 2-Hexanone-d5	8.030	63	35805	50.678	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.360%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	16145	5.081	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.600%	
66) 1,2-Dichlorobenzene-d4	11.561	152	27704	5.039	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.800%	
Target Compounds						
16) Methylene chloride	2.452	84	1164	0.252	ug/L	Qvalue 87
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV022024\  
 Data File : VV034278.D  
 Acq On : 20 Feb 2024 18:34  
 Operator : SY/MD  
 Sample : P1498-15 500X  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 COAH3

Quant Time: Feb 21 01:46:48 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR012524WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Feb 21 01:41:15 2024  
 Response via : Initial Calibration

