

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022122\
 Data File : VW024775.D
 Acq On : 21 Feb 2022 15:08
 Operator : SY/MD
 Sample : N1623-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBD25

Quant Time: Feb 22 03:29:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 22 03:27:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	169276	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	153440	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68785	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	48368	3.369	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	67.400%	
7) Chloroethane-d5	1.568	69	43768	3.716	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	74.400%	
11) 1,1-Dichloroethene-d2	2.108	63	65015	2.471	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	49.400%#	
20) 2-Butanone-d5	3.892	46	99752	48.640	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.280%	
24) Chloroform-d	4.349	84	88463	3.949	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	79.000%	
26) 1,2-Dichloroethane-d4	5.034	65	41693	4.175	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
32) Benzene-d6	5.050	84	182156	4.115	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.400%	
36) 1,2-Dichloropropane-d6	6.069	67	57321	4.400	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.000%	
41) Toluene-d8	7.317	98	154648	3.845	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	76.800%	
43) trans-1,3-Dichloroprop...	7.625	79	17120	3.945	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.000%	
46) 2-Hexanone-d5	8.088	63	74448	43.598	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.200%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33040	4.208	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	84.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	50090	4.209	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	84.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022122\
Data File : VV024775.D
Acq On : 21 Feb 2022 15:08
Operator : SY/MD
Sample : N1623-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBD25

Quant Time: Feb 22 03:29:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 22 03:27:30 2022
Response via : Initial Calibration

