

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022122\
 Data File : VW024777.D
 Acq On : 21 Feb 2022 15:55
 Operator : SY/MD
 Sample : N1623-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBD28

Quant Time: Feb 22 03:29:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 22 03:27:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	182112	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	169670	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73377	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	54057	3.500	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.000%
7) Chloroethane-d5	1.568	69	48817	3.852	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.000%
11) 1,1-Dichloroethene-d2	2.111	63	73760	2.606	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.200%#
20) 2-Butanone-d5	3.889	46	116470	52.789	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.580%
24) Chloroform-d	4.352	84	100545	4.172	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
26) 1,2-Dichloroethane-d4	5.034	65	48800	4.542	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
32) Benzene-d6	5.050	84	207702	4.243	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
36) 1,2-Dichloropropane-d6	6.069	67	65855	4.572	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.317	98	175115	3.937	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	7.625	79	19948	4.157	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.200%
46) 2-Hexanone-d5	8.088	63	88602	46.924	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.840%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39024	4.495	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	56036	4.414	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%
Target Compounds						
5) Vinyl chloride	1.310	62	1035	0.053	ug/L #	1
22) cis-1,2-Dichloroethene	3.918	96	6539	0.447	ug/L	95
25) Chloroform	4.378	83	3119	0.118	ug/L	82
34) Trichloroethene	5.921	95	13350	0.877	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022122\
Data File : VV024777.D
Acq On : 21 Feb 2022 15:55
Operator : SY/MD
Sample : N1623-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBD28

Quant Time: Feb 22 03:29:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 22 03:27:30 2022
Response via : Initial Calibration

