

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022122\
 Data File : VW024782.D
 Acq On : 21 Feb 2022 17:52
 Operator : SY/MD
 Sample : N1623-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBD31

Quant Time: Feb 22 03:31:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 22 03:27:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	162542	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	148315	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63805	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	47940	3.477	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.600%
7) Chloroethane-d5	1.571	69	43447	3.841	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.800%
11) 1,1-Dichloroethene-d2	2.111	63	65818	2.605	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.200%#
20) 2-Butanone-d5	3.889	46	100640	51.106	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.220%
24) Chloroform-d	4.349	84	91296	4.244	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
26) 1,2-Dichloroethane-d4	5.034	65	42382	4.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
32) Benzene-d6	5.053	84	185717	4.340	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.069	67	58701	4.662	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.317	98	155573	4.001	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.000%
43) trans-1,3-Dichloroprop...	7.625	79	17780	4.239	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.088	63	74899	45.378	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.760%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35458	4.672	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	51417	4.657	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022122\
Data File : VV024782.D
Acq On : 21 Feb 2022 17:52
Operator : SY/MD
Sample : N1623-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBD31

Quant Time: Feb 22 03:31:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 22 03:27:30 2022
Response via : Initial Calibration

