

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV022219\
 Data File : VV009336.D
 Acq On : 22 Feb 2019 18:32
 Operator : SY/MD
 Sample : K1579-06
 Misc : 6.77G/10ML/MSVOA_V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB603

Quant Time: Feb 23 01:08:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM022019S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 23 00:37:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	91450	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	88581	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	40668	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26804	22.03	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.12%
7) Chloroethane-d5	1.58	69	41927	29.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.52%
10) 1,1-Dichloroethene-d2	2.13	63	57018	19.06	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.24%
20) 2-Butanone-d5	3.95	46	18066	42.43	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.86%
24) Chloroform-d	4.40	84	51476	24.24	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.96%
26) 1,2-Dichloroethane-d4	5.08	65	29822	25.90	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.60%
29) Benzene-d6	5.10	84	105928	23.84	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.36%
33) 1,2-Dichloropropane-d6	6.12	67	34399	26.31	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.24%
37) Toluene-d8	7.36	98	98056	23.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.76%
38) trans-1,3-Dichloropropene-	7.67	79	14718	21.92	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.68%
39) 2-Hexanone-d5	8.14	63	15997	43.99	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.98%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	34344	24.03	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.12%
61) 1,2-Dichlorobenzene-d4	11.67	152	36894	24.70	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV022219\
Data File : VV009336.D
Acq On : 22 Feb 2019 18:32
Operator : SY/MD
Sample : K1579-06
Misc : 6.77G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB603

Quant Time: Feb 23 01:08:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM022019S.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 23 00:37:16 2019
Response via : Initial Calibration

