

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022222\
 Data File : VW024800.D
 Acq On : 22 Feb 2022 16:43
 Operator : SY/MD
 Sample : N1545-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 23 01:09:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 22 03:27:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	155900	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	142696	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62681	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	68060	5.147	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.000%
7) Chloroethane-d5	1.568	69	55580	5.123	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.400%
11) 1,1-Dichloroethene-d2	2.108	63	87324	3.604	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.000%
20) 2-Butanone-d5	3.896	46	101002	53.475	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.940%
24) Chloroform-d	4.349	84	103895	5.035	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.800%
26) 1,2-Dichloroethane-d4	5.034	65	48734	5.299	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.000%
32) Benzene-d6	5.050	84	222061	5.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.800%
36) 1,2-Dichloropropane-d6	6.069	67	65406	5.399	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.000%
41) Toluene-d8	7.317	98	190552	5.094	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.800%
43) trans-1,3-Dichloroprop...	7.625	79	20355	5.044	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.800%
46) 2-Hexanone-d5	8.088	63	74462	46.890	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.780%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	37685	5.161	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	56485	5.208	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022222\
Data File : VV024800.D
Acq On : 22 Feb 2022 16:43
Operator : SY/MD
Sample : N1545-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 23 01:09:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 22 03:27:30 2022
Response via : Initial Calibration

