


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022222\
 Data File : VV024789.D
 Acq On : 22 Feb 2022 09:25
 Operator : SY/MD
 Sample : BFB275
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB275

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Fri Feb 25 03:28:05 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	27403	PASS
75	95	30	80	51.1	74397	PASS
95	95	100	100	100.0	145472	PASS
96	95	5	9	6.7	9775	PASS
173	174	0.00	2	1.1	1116	PASS
174	95	50	120	68.6	99757	PASS
175	174	5	9	7.6	7594	PASS
176	174	95	105	97.2	96949	PASS
177	176	5	10	6.3	6060	PASS

m/z	Abundance
39.90	213.0
44.00	362.0

Instrument :
MSVOA_V
ClientSampleId :
BFB275

m/z	Abundance
35.90	1043.0
37.00	5760.0
38.00	5241.0
39.00	2300.0
40.00	277.0
43.90	962.0
45.00	993.0
46.00	141.0
46.90	1698.0
47.90	836.0
49.00	5533.0
50.00	27080.0
51.00	8286.0
52.00	334.0
54.90	333.0
55.90	1935.0
57.00	4142.0
57.90	153.0
59.90	1295.0
61.00	7090.0
62.00	6456.0
63.00	5315.0
64.00	513.0
64.90	413.0
67.00	416.0
68.00	14740.0
69.00	14897.0
70.00	1141.0
72.00	726.0
73.00	6571.0
74.00	25144.0
75.00	74160.0
76.00	6641.0
77.00	946.0
77.80	589.0
78.90	4563.0
79.90	1424.0
80.90	4941.0
81.90	1158.0
82.90	162.0
85.80	188.0
86.90	6466.0
87.90	6395.0
90.90	686.0
91.90	4470.0
93.00	6769.0
94.00	17504.0
95.00	145088.0
96.00	9523.0
97.00	281.0
103.80	734.0
104.90	229.0
105.80	595.0
107.00	139.0
110.70	126.0
111.70	110.0
112.80	139.0
114.90	205.0
115.80	575.0
116.90	904.0
117.90	566.0
118.90	725.0
127.80	559.0
128.90	264.0
129.90	494.0
130.80	305.0
134.90	220.0
136.90	248.0
139.80	120.0
140.90	1600.0
141.90	197.0
142.90	1538.0
144.90	197.0
145.80	231.0
147.80	350.0
148.70	121.0
150.00	138.0
151.90	101.0
152.70	109.0
154.80	319.0
156.80	164.0
158.80	120.0
160.80	173.0
170.80	155.0
172.00	646.0
173.00	1110.0
173.90	102840.0
175.00	7622.0
175.90	100168.0
177.00	6144.0
177.80	168.0

Instrument :
MSVOA_V
ClientSampleId :
BFB275

m/z	Abundance
36.00	1069.0
37.00	6103.0
38.00	5689.0
39.00	2566.0
40.00	251.0
43.90	921.0
45.00	1062.0
46.00	194.0
46.90	1720.0
47.90	979.0
49.00	5796.0
50.00	28544.0
51.00	8582.0
52.00	328.0
54.90	330.0
56.00	2168.0
57.00	4372.0
58.00	159.0
59.90	1402.0
61.00	7465.0
62.00	6826.0
63.00	5640.0
64.00	529.0
64.90	440.0
67.00	426.0
68.00	15288.0
69.00	15901.0
70.00	1208.0
72.00	825.0
73.00	6893.0
74.00	26744.0
75.00	77120.0
76.00	7108.0
77.00	953.0
77.80	635.0
78.90	4826.0
79.90	1579.0
80.90	5091.0
81.80	1169.0
82.90	159.0
85.70	184.0
86.90	6830.0
88.00	6794.0
90.90	706.0
92.00	4796.0
93.00	6970.0
94.00	18456.0
95.00	150848.0
96.00	10094.0
97.00	269.0
103.90	777.0
104.90	245.0
105.80	633.0
107.00	158.0
110.90	123.0
111.70	104.0
112.80	127.0
114.90	163.0
115.80	595.0
116.90	893.0
117.90	577.0
118.90	654.0
127.80	534.0
128.90	269.0
129.90	493.0
130.90	284.0
134.90	227.0
136.90	255.0
139.90	102.0
140.90	1606.0
141.90	222.0
142.90	1559.0
144.90	219.0
145.80	248.0
147.90	323.0
148.70	119.0
149.80	144.0
151.90	102.0
154.80	320.0
156.80	188.0
158.80	108.0
160.90	198.0
170.80	133.0
171.90	652.0
173.00	1154.0
173.90	103248.0
175.00	7919.0
175.90	100784.0
177.00	6297.0
177.90	164.0

Instrument :
MSVOA_V
ClientSampleId :
BFB275

m/z	Abundance
36.00	1028.0
37.00	5946.0
38.00	5497.0
39.00	2503.0
40.00	248.0
44.00	857.0
45.00	1057.0
45.90	201.0
47.00	1550.0
47.90	963.0
49.00	5529.0
50.00	26584.0
51.00	8107.0
51.90	299.0
54.90	302.0
56.00	2136.0
57.00	4042.0
58.10	160.0
59.90	1414.0
61.00	6913.0
62.00	6432.0
63.00	5508.0
64.00	528.0
65.00	423.0
66.00	108.0
66.90	369.0
68.00	14173.0
69.00	15149.0
70.00	1161.0
71.90	762.0
73.00	6614.0
74.00	25120.0
75.00	71912.0
76.00	6693.0
76.90	918.0
77.90	622.0
78.90	4645.0
79.90	1529.0
80.90	4755.0
81.80	1053.0
82.90	146.0
85.70	178.0
86.90	6508.0
88.00	6430.0
90.90	612.0
92.00	4542.0
93.00	6296.0
94.00	17336.0
95.00	140480.0
96.00	9707.0
97.00	231.0
103.80	695.0
104.90	237.0
105.80	599.0
107.00	153.0
110.90	117.0
112.80	110.0
114.90	118.0
115.80	530.0
116.90	838.0
117.90	563.0
118.90	609.0
127.90	486.0
128.90	233.0
129.90	444.0
130.90	225.0
134.90	239.0
136.90	254.0
140.00	103.0
140.90	1504.0
141.80	205.0
142.90	1429.0
144.90	208.0
145.80	260.0
147.90	293.0
148.80	101.0
149.80	130.0
154.80	279.0
156.80	206.0
158.90	109.0
160.90	202.0
170.80	111.0
171.90	627.0
173.00	1084.0
173.90	93184.0
175.00	7242.0
175.90	89896.0
177.00	5739.0
177.90	176.0

Instrument :
MSVOA_V
ClientSampleId :
BFB275