

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022322\
 Data File : VW024808.D
 Acq On : 23 Feb 2022 12:47
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV252

Quant Time: Feb 24 01:01:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 24 00:58:52 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	25034	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	24280	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	14227	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	15995	0.532	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 106.000%	
4) 1,2-Dichloroethane-d4	5.042	65	7141	0.528	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
7) 1,2-Dichloropropane-d6	6.068	67	10200	0.532	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 106.000%	
8) Toluene-d8	7.317	98	31026	0.542	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 108.000%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	6249	0.537	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	
Target Compounds						Qvalue
3) Vinyl chloride	1.310	62	15349	0.535	ug/L	99
6) Trichloroethene	5.925	95	10632	0.540	ug/L	99
9) 1,2-Dibromoethane	8.353	107	6243	0.550	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	4617	0.547	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.426	75	687	0.537	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
Data File : VV024808.D
Acq On : 23 Feb 2022 12:47
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV252

Quant Time: Feb 24 01:01:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 24 00:58:52 2022
Response via : Initial Calibration

