

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
 Data File : VV024810.D
 Acq On : 23 Feb 2022 14:40
 Operator : SY/MD
 Sample : N1331-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT1-2022-01

Quant Time: Feb 24 01:01:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 24 00:58:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	28156	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	26625	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	15595	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	17709	0.524	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 104.000%	
4) 1,2-Dichloroethane-d4	5.042	65	8117	0.534	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
7) 1,2-Dichloropropane-d6	6.068	67	11250	0.535	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 106.000%	
8) Toluene-d8	7.317	98	33569	0.534	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	6908	0.541	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	1040	0.032	ug/L #	59
6) Trichloroethene	5.925	95	684	0.032	ug/L	92
9) 1,2-Dibromoethane	8.368	107	397	0.032	ug/L #	90
12) 1,2,3-Trichloropropane	10.267	75	276	0.030	ug/L #	94
13) 1,2-Dibromo-3-chloropr...	12.426	75	42	0.030	ug/L #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
Data File : VV024810.D
Acq On : 23 Feb 2022 14:40
Operator : SY/MD
Sample : N1331-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT1-2022-01

Quant Time: Feb 24 01:01:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 24 00:58:52 2022
Response via : Initial Calibration

