

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
 Data File : VV024805.D
 Acq On : 23 Feb 2022 11:34
 Operator : SY/MD
 Sample : VSTD0.549
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5249

Quant Time: Feb 24 00:45:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 24 00:44:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	25827	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	25283	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	14801	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	15977	0.585	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	7105	0.425	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	10312	0.489	ug/L	0.00
8) Toluene-d8	7.317	98	30281	0.539	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	6272	0.492	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	15419	0.522	ug/L	100
6) Trichloroethene	5.925	95	10544	0.501	ug/L	100
9) 1,2-Dibromoethane	8.353	107	6227	0.495	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	4633	0.394	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.427	75	670	0.327	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
Data File : VV024805.D
Acq On : 23 Feb 2022 11:34
Operator : SY/MD
Sample : VSTD0.549
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5249

Quant Time: Feb 24 00:45:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 24 00:44:42 2022
Response via : Initial Calibration

