

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
 Data File : VV024807.D
 Acq On : 23 Feb 2022 12:21
 Operator : SY/MD
 Sample : VSTD00251
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002251

Quant Time: Feb 24 00:45:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 24 00:44:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	26126	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	25442	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	14926	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	64265	2.326	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.043	65	30297	1.792	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	41360	1.948	ug/L	0.00
8) Toluene-d8	7.317	98	128489	2.272	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.199	84	25236	1.967	ug/L	-0.02
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	59889	2.005	ug/L	98
6) Trichloroethene	5.925	95	42903	2.025	ug/L	98
9) 1,2-Dibromoethane	8.353	107	24542	1.937	ug/L #	97
12) 1,2,3-Trichloropropane	10.267	75	18161	1.533	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.410	75	2571	1.243	ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022322\
Data File : VV024807.D
Acq On : 23 Feb 2022 12:21
Operator : SY/MD
Sample : VSTD00251
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002251

Quant Time: Feb 24 00:45:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 24 00:44:42 2022
Response via : Initial Calibration

