



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022422\
 Data File : VV024812.D
 Acq On : 24 Feb 2022 10:01
 Operator : SY/MD
 Sample : BFB276
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB276

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Fri Feb 25 03:28:05 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	28907	PASS
75	95	30	80	51.6	83755	PASS
95	95	100	100	100.0	162176	PASS
96	95	5	9	6.5	10588	PASS
173	174	0.00	2	1.0	1173	PASS
174	95	50	120	70.3	114072	PASS
175	174	5	9	7.3	8375	PASS
176	174	95	105	95.1	108445	PASS
177	176	5	10	6.2	6701	PASS

m/z	Abundance
39.90	273.0
43.90	341.0

Instrument :
MSVOA_V
ClientSampleId :
BFB276

m/z	Abundance
36.00	1312.0
37.00	7049.0
38.00	6274.0
39.00	2731.0
39.90	432.0
42.80	150.0
43.90	1216.0
44.90	1383.0
46.00	202.0
47.00	2271.0
48.00	782.0
49.00	6278.0
50.00	29744.0
51.00	9370.0
52.00	434.0
55.00	416.0
56.00	2290.0
57.00	4452.0
57.90	155.0
60.00	1562.0
61.00	7798.0
62.00	7813.0
63.00	6471.0
64.00	533.0
65.00	528.0
66.90	477.0
68.00	16832.0
69.00	17280.0
70.00	1274.0
71.90	901.0
73.00	7453.0
74.00	29736.0
75.00	86592.0
76.00	7221.0
77.00	1048.0
77.90	708.0
78.90	5444.0
79.90	1601.0
80.90	5494.0
81.90	1370.0
82.80	200.0
86.00	297.0
87.00	7937.0
87.90	7918.0
90.90	703.0
92.00	5114.0
93.00	7501.0
94.00	20472.0
95.00	167936.0
96.00	10931.0
97.00	304.0
102.70	117.0
103.90	920.0
104.90	278.0
105.80	799.0
106.80	187.0
109.80	131.0
110.80	136.0
112.80	218.0
114.80	243.0
115.90	658.0
116.90	980.0
117.80	685.0
118.80	853.0
127.90	555.0
128.80	224.0
129.90	659.0
130.80	242.0
134.90	299.0
136.80	306.0
140.00	104.0
140.80	1669.0
141.80	214.0
142.90	1746.0
143.80	114.0
144.90	247.0
145.90	204.0
147.00	120.0
147.80	308.0
149.80	170.0
152.90	131.0
153.80	106.0
154.90	365.0
156.80	250.0
158.80	128.0
160.90	213.0
172.00	1005.0
172.90	1223.0
173.90	117928.0
175.00	8616.0
175.90	112480.0
176.90	6872.0

Instrument :
MSVOA_V
ClientSampleId :
BFB276

177.90

165.0

Instrument :
MSVOA_V
ClientSampleId :
BFB276

m/z	Abundance
35.90	1254.0
37.00	6496.0
38.00	5915.0
39.00	2427.0
39.90	396.0
42.80	137.0
43.90	1161.0
44.90	1351.0
46.00	208.0
47.00	2210.0
47.90	770.0
49.00	6186.0
50.00	28528.0
51.00	8772.0
52.00	340.0
55.00	357.0
56.00	2309.0
57.00	4208.0
57.90	160.0
60.00	1448.0
61.00	7299.0
62.00	7380.0
63.00	6128.0
64.00	541.0
65.00	518.0
66.90	398.0
68.00	16051.0
69.00	16608.0
70.00	1163.0
72.00	854.0
73.00	7334.0
74.00	28712.0
75.00	83728.0
76.00	6632.0
77.00	943.0
78.00	763.0
78.90	5126.0
79.90	1652.0
80.90	5205.0
81.90	1360.0
82.80	162.0
86.00	308.0
87.00	7522.0
87.90	7604.0
90.90	693.0
92.00	4925.0
93.00	7006.0
94.00	19800.0
95.00	162176.0
96.00	10430.0
97.00	303.0
102.70	121.0
103.80	910.0
104.90	283.0
105.80	846.0
106.80	209.0
109.80	104.0
110.80	125.0
112.80	179.0
114.80	224.0
115.90	586.0
116.90	999.0
117.80	700.0
118.80	835.0
127.90	536.0
128.80	248.0
129.80	574.0
130.80	224.0
134.90	289.0
136.80	284.0
140.00	113.0
140.90	1648.0
141.90	254.0
142.90	1673.0
144.90	209.0
145.90	227.0
147.00	106.0
147.80	314.0
149.90	142.0
152.90	101.0
155.00	347.0
156.80	270.0
158.90	150.0
160.90	236.0
172.00	916.0
172.90	1170.0
173.90	117432.0
175.00	8559.0
175.90	111976.0
176.90	6835.0
177.90	187.0

Instrument :
MSVOA_V
ClientSampleId :
BFB276

m/z	Abundance
36.00	1278.0
37.00	6808.0
38.00	5971.0
39.00	2740.0
39.90	418.0
42.80	121.0
43.90	1119.0
44.90	1307.0
46.00	168.0
47.00	2075.0
48.00	761.0
49.00	6090.0
50.00	28448.0
51.00	9210.0
52.00	500.0
55.00	437.0
55.90	2079.0
57.00	4138.0
57.90	130.0
59.90	1488.0
61.00	7525.0
62.00	7261.0
63.00	6049.0
64.00	580.0
64.90	532.0
67.00	467.0
68.00	15807.0
69.00	16241.0
70.00	1225.0
71.90	852.0
73.00	7069.0
74.00	27784.0
75.00	80944.0
76.00	7070.0
76.90	1090.0
77.90	613.0
78.90	5163.0
79.90	1411.0
80.90	5259.0
81.90	1243.0
82.80	195.0
85.90	240.0
87.00	7478.0
87.90	7208.0
90.90	651.0
92.00	4667.0
93.00	7068.0
94.00	19008.0
95.00	156416.0
96.00	10404.0
97.00	243.0
102.80	109.0
103.90	805.0
104.90	235.0
105.80	650.0
106.80	170.0
109.80	109.0
110.70	126.0
112.80	195.0
114.80	212.0
115.90	682.0
116.90	931.0
117.80	616.0
118.80	799.0
123.80	113.0
127.80	524.0
128.90	203.0
129.90	649.0
130.80	250.0
134.90	242.0
136.80	299.0
140.80	1539.0
141.80	164.0
142.90	1600.0
143.80	153.0
144.90	227.0
145.90	184.0
146.90	109.0
147.80	302.0
149.80	182.0
152.80	126.0
154.80	348.0
156.80	212.0
158.80	131.0
160.90	185.0
171.00	132.0
171.90	926.0
172.90	1126.0
173.90	106856.0
175.00	7950.0
175.90	100880.0
176.90	6396.0

Instrument :
MSVOA_V
ClientSampleId :
BFB276

177.90

138.0

Instrument :
MSVOA_V
ClientSampleId :
BFB276