

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022522\
 Data File : VV024824.D
 Acq On : 25 Feb 2022 10:11
 Operator : SY/MD
 Sample : VV0225WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Feb 28 00:47:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Feb 28 00:47:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	32840	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.848	117	30744	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	17826	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	17194	0.436	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	88.000%
4) 1,2-Dichloroethane-d4	5.042	65	8066	0.455	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
7) 1,2-Dichloropropane-d6	6.068	67	11351	0.467	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	94.000%
8) Toluene-d8	7.317	98	34406	0.474	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	7120	0.483	ug/L	0.02
Spiked Amount	0.500	Range	65 - 120	Recovery	=	96.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022522\
Data File : VV024824.D
Acq On : 25 Feb 2022 10:11
Operator : SY/MD
Sample : VV0225WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Feb 28 00:47:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM022322.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Feb 28 00:47:14 2022
Response via : Initial Calibration

