

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022724\
 Data File : VW034439.D
 Acq On : 27 Feb 2024 11:11
 Operator : SY/MD
 Sample : P1535-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF9

Quant Time: Feb 28 03:54:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 28 03:52:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	147615	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	139550	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	70871	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	37101	4.330	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.600%
7) Chloroethane-d5	1.536	69	33687	4.451	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.060	65	17764	4.454	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.000%
20) 2-Butanone-d5	3.806	46	116738	52.106	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.220%
24) Chloroform-d	4.252	84	80568	4.683	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
26) 1,2-Dichloroethane-d4	4.947	65	42690	4.876	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	4.963	84	142533	4.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
36) 1,2-Dichloropropane-d6	5.992	67	44489	4.746	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.246	98	125070	4.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%
43) trans-1,3-Dichloroprop...	7.564	79	15712	4.499	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.027	63	104915	53.134	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.260%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	35676	4.809	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	52282	4.856	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.069	101	5751	0.599	ug/L	92
19) 1,1-Dichloroethane	3.114	63	29289	1.619	ug/L	98
22) cis-1,2-Dichloroethene	3.818	96	30004	2.902	ug/L	97
25) Chloroform	4.285	83	25365	1.350	ug/L	99
29) 1,1,1-Trichloroethane	4.519	97	11628	0.666	ug/L	100
31) Carbon tetrachloride	4.738	117	21881	1.444	ug/L	100
34) Trichloroethene	5.841	95	15011	1.488	ug/L	94
47) Tetrachloroethene	7.905	164	65957	8.336	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022724\
Data File : VV034439.D
Acq On : 27 Feb 2024 11:11
Operator : SY/MD
Sample : P1535-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF9

Quant Time: Feb 28 03:54:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 28 03:52:36 2024
Response via : Initial Calibration

