

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022724\
 Data File : VW034450.D
 Acq On : 27 Feb 2024 15:35
 Operator : SY/MD
 Sample : P1566-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JWAD2

Quant Time: Feb 28 03:57:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 28 03:52:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	136318	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	131624	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	67811	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	33918	4.286	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.800%
7) Chloroethane-d5	1.532	69	30755	4.401	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	2.060	65	15687	4.259	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.200%
20) 2-Butanone-d5	3.805	46	112110	54.187	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.380%
24) Chloroform-d	4.252	84	76531	4.817	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	4.947	65	41881	5.180	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
32) Benzene-d6	4.963	84	135409	4.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	5.992	67	43794	4.953	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.246	98	117355	4.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
43) trans-1,3-Dichloroprop...	7.561	79	15185	4.610	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.200%
46) 2-Hexanone-d5	8.027	63	101113	54.292	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.580%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	35841	5.122	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.400%
66) 1,2-Dichlorobenzene-d4	11.561	152	50663	4.918	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.400%
Target Compounds						Qvalue
13) Acetone	2.134	43	17330	8.236	ug/L	99
25) Chloroform	4.285	83	4860	0.280	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022724\
Data File : VV034450.D
Acq On : 27 Feb 2024 15:35
Operator : SY/MD
Sample : P1566-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JWAD2

Quant Time: Feb 28 03:57:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 28 03:52:36 2024
Response via : Initial Calibration





#13

Acetone

Concen: 8.236 ug/L

RT: 2.134 min Scan# 34

Delta R.T. -0.010 min

Lab File: VV034450.D

Acq: 27 Feb 2024 15:35

Instrument :

MSVOA_V

ClientSampleId :





JWAD2

Tgt Ion: 43 Resp: 17330

Ion Ratio Lower Upper

43 100

58 27.4 0.0 56.2



#25

Chloroform

Concen: 0.280 ug/L

RT: 4.285 min Scan# 1009

Delta R.T. 0.010 min

Lab File: VV034450.D

Acq: 27 Feb 2024 15:35

Tgt Ion: 83 Resp: 4860

Ion Ratio Lower Upper

83 100

85 63.4 47.0 87.4

