


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030619\
 Data File : VV009509.D
 Acq On : 06 Mar 2019 09:23
 Operator : SY/MD
 Sample : BFB70
 Misc : 5.00G/10ML/MSVOA_V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB70

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
 Title : VOC Analysis
 Last Update : Fri Mar 08 03:01:51 2019



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2795

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	13515	PASS
75	95	30	80	55.2	37842	PASS
95	95	100	100	100.0	68506	PASS
96	95	5	9	6.6	4547	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.0	51352	PASS
175	174	5	9	8.0	4100	PASS
176	174	95	101	97.2	49933	PASS
177	176	5	9	6.7	3370	PASS

m/z	Abundance
35.90	556.0
37.00	3424.0
38.00	3279.0
39.00	1407.0
39.90	306.0
43.00	132.0
43.90	784.0
44.90	804.0
46.20	155.0
47.00	770.0
48.00	404.0
49.00	3131.0
50.00	14196.0
51.00	4049.0
52.00	195.0
54.80	219.0
55.10	227.0
56.10	1518.0
57.00	2459.0
58.00	158.0
59.90	672.0
61.00	3599.0
62.00	3835.0
63.00	2834.0
64.10	381.0
64.90	179.0
67.00	239.0
68.00	8158.0
69.00	8515.0
70.00	717.0
71.90	512.0
73.00	3832.0
74.00	14127.0
75.00	38816.0
76.00	3271.0
77.10	406.0
77.80	272.0
78.90	3610.0
79.90	1064.0
80.90	4082.0
81.90	890.0
83.00	124.0
86.90	2646.0
88.00	2470.0
90.90	415.0
92.00	2270.0
93.00	3872.0
94.00	9244.0
95.00	70360.0
96.00	4804.0
96.90	138.0
103.90	407.0
104.90	176.0
105.90	382.0
107.00	118.0
110.70	111.0
113.00	168.0
114.60	118.0
115.90	327.0
116.90	488.0
117.80	369.0
118.90	407.0
127.80	244.0
128.90	165.0
130.00	321.0
130.90	122.0
134.70	126.0
136.70	163.0
140.80	1011.0
142.90	1186.0
144.60	108.0
147.80	172.0
154.90	149.0
156.80	127.0
171.90	232.0
174.00	51816.0
175.00	4239.0
176.00	49528.0
177.00	3333.0
178.00	113.0

Instrument :
MSVOA_V
ClientSampleId :
BFB70

m/z	Abundance
35.90	575.0
37.00	3288.0
38.00	3121.0
39.00	1267.0
39.90	269.0
43.00	142.0
43.90	674.0
44.90	823.0
45.90	104.0
47.00	620.0
48.00	356.0
48.90	3043.0
50.00	13865.0
51.00	4196.0
52.00	206.0
54.70	229.0
56.00	1414.0
57.10	2438.0
58.10	184.0
59.90	711.0
60.90	3686.0
62.00	3792.0
63.00	3026.0
64.00	383.0
64.90	134.0
67.00	207.0
68.00	8104.0
69.00	8619.0
70.00	671.0
71.90	471.0
73.00	3750.0
74.00	14045.0
75.00	38944.0
76.00	3311.0
76.90	386.0
77.80	355.0
78.90	3703.0
79.90	1166.0
80.90	4105.0
81.90	868.0
86.90	2561.0
87.90	2491.0
90.90	448.0
92.00	2254.0
93.00	3769.0
94.00	9654.0
95.00	70608.0
96.00	4749.0
96.90	158.0
104.00	412.0
104.90	140.0
105.80	362.0
106.90	129.0
110.80	131.0
113.00	164.0
115.10	116.0
115.90	282.0
116.90	477.0
117.80	362.0
118.90	391.0
127.80	213.0
128.90	148.0
129.90	375.0
130.90	125.0
134.70	135.0
136.80	130.0
140.90	1089.0
142.90	1130.0
145.70	112.0
147.80	232.0
154.90	137.0
156.70	160.0
171.90	260.0
174.00	52880.0
175.00	4181.0
176.00	51272.0
176.90	3526.0

Instrument :
MSVOA_V
ClientSampleId :
BFB70

m/z	Abundance
36.00	534.0
37.00	2941.0
38.00	2819.0
39.00	1119.0
40.00	237.0
43.90	622.0
44.90	809.0
45.80	126.0
46.90	520.0
47.80	297.0
48.10	295.0
48.90	2832.0
50.00	12485.0
51.00	3919.0
52.00	196.0
54.70	207.0
56.00	1219.0
57.10	2160.0
58.10	163.0
59.90	713.0
60.90	3534.0
62.00	3294.0
63.00	2891.0
64.00	322.0
66.90	175.0
68.00	7232.0
69.00	7960.0
70.00	608.0
71.90	340.0
73.00	3354.0
74.00	12602.0
75.00	35768.0
76.00	2968.0
76.90	329.0
77.80	410.0
78.90	3467.0
79.90	1106.0
80.90	3617.0
81.90	761.0
86.90	2394.0
87.90	2401.0
90.90	376.0
92.00	2136.0
93.00	3376.0
94.00	9101.0
95.00	64552.0
96.00	4090.0
96.90	136.0
104.00	438.0
104.90	107.0
105.80	352.0
106.80	141.0
110.80	147.0
113.00	104.0
115.20	142.0
115.90	268.0
117.00	456.0
117.70	330.0
119.00	365.0
127.80	231.0
128.90	116.0
129.90	418.0
130.80	126.0
134.60	123.0
137.00	124.0
140.90	1023.0
142.90	969.0
145.70	122.0
147.80	251.0
154.90	125.0
156.70	174.0
171.10	130.0
171.90	246.0
174.00	49360.0
175.00	3882.0
176.00	49000.0
176.90	3252.0

Instrument :
MSVOA_V
ClientSampleId :
BFB70

m/z	Abundance
39.80	375.0
44.00	152.0

Instrument :
MSVOA_V
ClientSampleId :
BFB70