

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030822\
 Data File : VW024868.D
 Acq On : 08 Mar 2022 13:29
 Operator : SY/MD
 Sample : N1811-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDA9

Quant Time: Mar 09 05:03:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 09 05:01:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	177864	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	165620	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	70969	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69322	4.595	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.568	69	61207	4.945	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.108	63	90741	3.282	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	3.899	46	132002	61.257	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.520%
24) Chloroform-d	4.349	84	129475	5.500	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.000%
26) 1,2-Dichloroethane-d4	5.034	65	56885	5.421	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.400%
32) Benzene-d6	5.050	84	270500	5.661	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.200%
36) 1,2-Dichloropropane-d6	6.069	67	82211	5.847	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.000%
41) Toluene-d8	7.317	98	229039	5.276	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
43) trans-1,3-Dichloroprop...	7.622	79	26353	5.626	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	112.600%
46) 2-Hexanone-d5	8.088	63	99195	53.819	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.640%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	46768	5.519	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	70056	5.705	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.200%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.121	96	2612	0.189	ug/L	# 1
18) trans-1,2-Dichloroethene	2.770	96	584	0.043	ug/L	97
22) cis-1,2-Dichloroethene	3.918	96	15034	1.051	ug/L	96
25) Chloroform	4.372	83	2088	0.081	ug/L	76
34) Trichloroethene	5.912	95	800294	53.877	ug/L	96
47) Tetrachloroethene	7.976	164	862	0.085	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030822\
Data File : VV024868.D
Acq On : 08 Mar 2022 13:29
Operator : SY/MD
Sample : N1811-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDA9

Quant Time: Mar 09 05:03:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR021722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 09 05:01:56 2022
Response via : Initial Calibration

