

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031022\
 Data File : VW024914.D
 Acq On : 10 Mar 2022 14:33
 Operator : SY/MD
 Sample : N1855-15
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D43

Quant Time: Mar 11 02:48:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 02:45:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	294601	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	284929	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	127440	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	127183	42.457	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.920%
7) Chloroethane-d5	1.568	69	98288	41.821	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.640%
11) 1,1-Dichloroethene-d2	2.108	63	160571	31.937	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.880%
21) 2-Butanone-d5	3.896	46	140165	82.220	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.220%
24) Chloroform-d	4.349	84	208391	39.852	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.700%
26) 1,2-Dichloroethane-d4	5.031	65	126651	41.050	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.100%
32) Benzene-d6	5.050	84	446918	42.318	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.640%
36) 1,2-Dichloropropane-d6	6.069	67	139388	42.362	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.720%
41) Toluene-d8	7.313	98	389081	41.585	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.180%
43) trans-1,3-Dichloroprop...	7.622	79	56483	39.506	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.020%
47) 2-Hexanone-d5	8.088	63	94021	75.280	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.280%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	161311	41.333	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.660%
66) 1,2-Dichlorobenzene-d4	11.622	152	129817	44.221	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024914.D
Acq On : 10 Mar 2022 14:33
Operator : SY/MD
Sample : N1855-15
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D43

Quant Time: Mar 11 02:48:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 02:45:24 2022
Response via : Initial Calibration

